

Certificate of Analysis

Date:	2025-01-19 14:31:13 -06:00	Cultivar:	Cereal Milk
Serial:	LL0339	Moisture:	0.0%
LightLab	LAB	Col Tests Remaining:	22
Operator:	MF	CoA Revision:	0
Sample ID:	F353	Calibration Exp:	2025-03-14
Method:	LightLab HPLC	Product:	Deep Gap Wellness
Test Type:	THCA Flower/Sprayed Flower	SKU:	THC-A
Weight / Volume:	0.36 g	Batch:	
Solvent:	20 ml		
Temperature:	21.6 °C		


Cannabinoid Profile

Notes:

Analyte		%	LOQ
THC-A		35.0	0.28
Δ9-THC		ND	0.22
CBD-A		ND	0.83
CBG-A		ND	0.83
CBD/CBG		ND	0.28
CBN		ND	0.83
CBC-A		1.7	0.28
CBC		ND	0.83
∆8-THC*		ND	2.2
Δ10-THC		ND	0.83
THCV-A		ND	0.83
THCV		ND	0.83
Δ9-THC-O		ND	0.83
∆8-THC-O		ND	0.83
HHC		ND	0.83
Δ9-THCP		ND	0.83
Terpenes		Low	
Total THC		30.7	
Total Cannal	binoids	36.7	

ND = Not Detected; n/a = Not Analyzed; LOQ = Limit of Quantification; Total THC = (0.877 x THC-A) + Δ 9-THC; Total CBD = (0.877 x CBD-A) + CBD. * Δ 8THC has lower precision and higher detection limit than other cannabinoids.

Chromatogram.



Change History

2025-01-19 14:31:13 -06:00 Test Recorded 2025-01-23 12:38:34 -06:00 Generated a CoA (revision 0) Image: Comparison of the second description of	Date	User	Action	
2025-01-23 12:38:34 -06:00 Generated a CoA (revision 0)		0301		
Scan for Authenticity	2025-01-23 12:38:34 -06:00		Generated a CoA (revision 0)	
Scan for Authenticity				
Scan for Authenticity				
Scan for Authenticity				
Scan for Authenticity				in second in the second s
Scan for Authenticity				
Scan for Authenticity	\mathcal{Y}			
	flo		Jan 23, 2025	
	Approved		Date	Scan for Authenticity

The signatory confirms that the Operator has performed the sample preparation according to the LightLab User's Guide. This report is for quality assurance purposes only. These results relate only to the sample included on this report. Orange Photonics makes no claims as to the efficacy, safety, or risks associated with any detected or non-detected level of any compounds reported herein. Orange Photonics makes no claims regarding the adherence to sample preparation guidelines, by the operator, as outlined in the LightLab User's Guide.